

Overview of the ACMCC particulate organonitrates (pON) intercomparison

Alexandre Albinet¹, Jean-Eudes Petit², Andrew Lambe³, Athina Kalogridis⁴, Liine Heikkinen⁵, Frans Graeffe⁵, Manuela Cirtog⁶, Anaïs Féron⁶, James Allan⁷, Zainab Bibi⁷, Tanguy Amodeo¹, Nicolas Karoski¹, Robin Aujay-Plouzeau¹, Laurent Meunier¹, Camille Noblet^{1,8}, François Lestremau¹, Jean-Luc Besombes⁸, Valerie Gros², Nicolas Bonnaire², Roland Sarda-Estève², François Truong², Mikael Ehn⁵, Tuija Jokinen⁵, Minna Aurela⁹, Marek Maasikmets¹⁰, Christina Marin¹¹, Luminita Marmureanu¹¹, Axel Eriksson¹², Erik Ahlberg¹², Evelyn Freney¹³, MariCruz Minguillon¹⁴, Philip Croteau³, John Jayne³, Leah Williams³, Olivier Favez¹

¹ Institut National de l'Environnement Industriel et des Risques (INERIS), Verneuil-en-Halatte, France

² Laboratoire des Sciences du Climat et de l'Environnement (LSCE), Gif-sur-Yvette, France

³ Aerodyne Research, Inc. (ARI), Billerica, Massachusetts, USA

⁴ NCSR Demokritos, Athens, Greece

⁵ Institute for Atmospheric and Earth System Research (INAR), University of Helsinki, Finland

⁶ Laboratoire Universitaire des Systèmes Atmosphériques (LISA), Créteil, France

⁷ School of Earth and Environmental Sciences, The University of Manchester, UK

⁸ Université Savoie Mont Blanc, Le Bourget du Lac, France

⁹ Finnish meteorological institute (FMI), Helsinki, Finland

¹⁰ Estonian Environmental Research Center (EERC), Tallinn, Estonia

¹¹ National Institute of R&D for Optoelectronics (INOE), Ilfov, Romania

¹² Lund University, Nuclear physics, Lund, Sweden

¹³ Laboratoire de Météorologie Physique (LaMP), Aubière, France

¹⁴ Institute of Environmental Assessment and Water Research (IDAEA-CSIC), Barcelona, Spain

Keywords: Organonitrates; Secondary organic aerosol (SOA); ACSM; Oxidation flow reactor (OFR).

Contact: alexandre.albinet@ineris.fr

Introduction

Particulate organonitrates (pON) account for significant fraction (5-80% by mass) of total OA in ambient air. They are formed from the reactions of volatile organic compounds (VOCs) with atmospheric oxidants (OH/NO₃ radicals) and NO_x. Their quantification can be achieved using aerosol mass spectrometry (AMS), based on the characteristic mass fragment ratio (NO₂⁺/NO⁺) allowing the distinction from inorganic nitrate. However, the accuracy of the low-resolution aerosol chemical speciation monitor (ACSM) to determine pON has not yet been evaluated. At the Aerosol Chemical Monitor Calibration Centre (AMCC), an intercomparison for the measurements of pON has been performed in order to obtain a stable and constant generation of pON, so to compare simultaneously the response of nine different AMS/ACSM systems (long-TOF-AMS vs ACSMs; Quads vs TOFs; standard vs capture vaporizers), as well as to investigate the pON physical properties and chemical composition.

Methods

pON were generated in a Potential Aerosol Mass (PAM) oxidation flow reactor from the reaction of NO₃ radical, produced on demand (O₃ + NO₂), with single VOC precursors. Two biogenic (limonene and β-pinene) and two anthropogenic (acenaphthylene and guaiacol) pON precursors were investigated. For the determination of AMS/ACSM relative ionization efficiencies (RIE), a particle size and mass selection were achieved by combining an

aerodynamic aerosol classifier (AAC) and centrifugal a particle mass analyser (CPMA). pON size distribution and total particle number concentration were monitored by a scanning mobility particle sizer (SMPS) and a condensation particle counter (CPC) allowing the characterization of the pON density (Figure 1). In order to get insights into the pON optical properties, as well as their chemical composition and formation processes, measurements also included cavity-enhanced absorption spectroscopy (NO₃ radical by IBB-CEAS), proton-transfer-reaction MS (PTR-MS), multi-wavelengths aethalometer (AE33), as well as filter samplings for further high-resolution MS off line analyses (GC and LC/Q-TOF-MS).

An overview of the set-up and the experiments performed will be presented together with preliminary key results.




Figure 1. Evaluation of the pON density.

This work is part of the COST Action CA16109 COLOSSAL and the Aerosol, Clouds, and Trace gases Research InfraStructure (ACTRIS).